

भोपाल स्मारक अस्पताल एवं अनुसंधान केंद्र
BHOPAL MEMORIAL HOSPITAL & RESEARCH CENTRE

BID Ref. No BMHRC/PUR/Oct/2018-19/66

Bhopal Memorial Hospital & Research Centre is a 350 bedded Super Specialty Hospital situated at Bhopal. The Hospital and its eight Mini Units is providing free treatment to the Gas Victims and their dependents of the Bhopal Gas Tragedy, free of cost. The Hospital is being run by the Department of Health Research, Ministry of Health and Family Welfare, Govt. of India.

Sealed bid are hereby invited on behalf of Director BMHRC from the intending registered firms Authorized Suppliers for the following item's in the prescribed format attached to this mail:-

S.No.	Item Name	Qty.
01	<p>T3 FREE TEST KIT</p> <p>(Manual ELISA kit for Free T3 estimation.)</p> <p style="padding-left: 40px;">High control and low control should also be provided.</p> <ul style="list-style-type: none"> • Total test procedure must complete in less than 100 minutes. • Reading should be taken at 450 nm with reference option of 620 – 630 nm • Material should include - <ul style="list-style-type: none"> - Coated wells 96 wells pre-coated with monoclonal anti-Free T3. - Calibration - 6 to 7 calibrators. - Enzyme conjugate reagent HRP labeled monoclonal Anti-Free T3 in Tris-NaCl buffer. - Concentrated washing solution - TMB substrate - Stop solution. - Kit should be approved by USFDA / European CE for IVD purpose. - Demonstration of kit should be done by company technical person and results of kits should also be validated with external quality control samples before approval. . 	10 Kits
02	<p>T4 FREE TEST KIT</p> <ul style="list-style-type: none"> • Manual ELISA kit for Free T4 estimation. • High control and low control should also be provided. • Total test procedure must complete in less than 100 minutes. • Reading should be taken at 450 nm with reference option of 620 – 630 nm • Material should include - <ul style="list-style-type: none"> - Coated wells 96 wells pre-coated with monoclonal anti-Free T4. - Calibration - 6 to 7 calibrators. - Enzyme conjugate reagent HRP labeled monoclonal Anti-Free T4 in Tris-NaCl buffer. - Concentrated washing solution - TMB substrate - Stop solution. - Kit should be approved by USFDA / European CE for IVD purpose. - Demonstration of kit should be done by company technical person and results of kits should also be validated with external quality control samples before approval. 	10 kits

03	<p>TSH TEST KIT</p> <ul style="list-style-type: none"> - Manual ELISA kit for Total TSH estimation. - High control and low control should also be provided 1. Total test procedure must complete in less than 100 minutes. 2. Reading should be taken at 450 nm with reference option of 620 – 630 nm i. Material should include - <ul style="list-style-type: none"> - Coated wells 96 wells pre-coated with monoclonal anti-TSH. - Calibration - 6 to 7 calibrators. - Enzyme conjugate reagent HRP labeled monoclonal Anti-TSH in Tris-NaCl buffer. - Concentrated washing solution - TMB substrate - Stop solution. - Kit should be approved by USFDA / European CE for IVD purpose. - Demonstration of kit should be done by company technical person and results of kits should also be validated with external quality control samples before approval. 	10 kits
----	---	---------

1. Please send the bids with best possible rates on FOR basis.
2. Envelope should be clearly marked as "**Offer for No. BMHRC /PUR /Oct/2018-19/66**
3. The sealed bid should reach in this office by 11.00 AM on or before **17.10.2018** Bid send by E-mail / Fax will not be considered.
4. Bid must be send in sealed envelope by hand/by post or by courier before the deadline for submission of bids at **17.10.2018** or bids may be dropped in “ Quotation Box” kept in basement of Administration Block (Near Deptt. Of Purchase) BMHRC.
5. Authorization Certificate for supply of the item is mandatory,
6. Full specifications of items being offered and illustrated literature are to accompany with the bid.
7. Bidders should ensure that their bid complies to the indicated requirements/specification. Yes should be marked in the ‘Compliance column”(if applicable).
8. BMHRC reserves the right to reject any or all bids without assigning any reason thereof.
9. Kindly Provide the Bid in prescribed format only.
10. Offer will be valid for one year from the opening of the bid. The quoted rates should be kept open/valid for a period of one year from the date of signing the agreement.

Chairman
Local Purchase Committee
Bhopal Memorial Hospital & Research Centre
Raisen By Pass Road, Karond
Bhopal - 462 038

**To
Director BMHRC
Bhopal**

Sub: Offer for quotation no. BMHRC/PUR/Oct/2018-19/66
With reference to above, our offer is as under:-

Sl. No.	Item Description / Cat No.	Req. Qty.	Pack Size	Company/ Brand Name	MRP	List Price	Discount in %	GST in %	Price to BMHRC (Including GST & Discount)	Total Net Price
1	2	3	4	5	6	7	8	9	10	11
01	T3 FREE TEST KIT	10kits								
02	T4 FREE TEST KIT	10 kits								
03	TSH TEST KIT	10kits								

I/We, the undersigned, hereby declare that

I/We certify that quoted product meets the specification as described in the call for bid.

I/ We certify that rate quoted for this product does not exceed the Maximum Retail Price of the product.

I/We agree, in case of expiry/near expiry of supplied goods against this bid bidder should replace the said goods with fresh batch without any financial implication. (If applicable).

Name and Surname:

Signature of Bidder:

Stamp of the firm/company