

भोपाल स्मारक अस्पताल एवं अनुसंधान केन्द्र

BHOPAL MEMORIAL HOSPITAL & RESEARCH CENTRE

स्वास्थ्यअनुसंधान विभाग, स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार
 Department Of Health Research, Ministry Of Health And Family Welfare, Government Of India
 रायसेन बायपास रोड, भोपाल - 462038 (म. प्र.) Raisen Bypass Road Bhopal - 462038 (M.P.)
 Phones: +91 755 2742212-16, Fax: - +91 755 2748309, Email: bmhrcbhopal@gmail.com Website: WWW.bmhrc.org

CORRIGENDUM NOTICE

Subject: Corrigendum Notice for Tender No BMHRC/PUR/RE/2016-17/06 (Modular Monitor with Central Station).

The following amendments in Technical Specification are approved by the Technical Specification committee:

S. No.	Particular /Technical Specification	Consideration of Committee Members
1	Point 4: Minimum 17 inches multicolored TFT display screen operation through knob and touch screen (optional).	Read as Point 4: Minimum 19 to 21 inches multicolored TFT display screen operation through knob and touch screen (optional).
2	Point 6: Eight digital and waveforms /traces display	Read as Point 6: Minimum Eight digital and waveforms /traces display
3	Point 10: Facility to monitor and display - ECG, respiration, NIBP, IBP, Spo2 CO2 with capnography, Temp. Cardiac output, NMT, Optional -BIS/Entropy	Read as Point 10: Facility to monitor and display - ECG, respiration, NIBP, IBP-2, Spo2 CO2 with capnography, Temp. Cardiac output, NMT, Optional -BIS/Entropy
4	Point 17: 200 nos. events recall / snapshot facility both manually and automatically triggered by alarm	Read as Point 17: Minimum 150 nos. events recall / snapshot facility both manually and automatically triggered by alarm
5	Point 21: Battery backup of upto 2hours when charged, in all high end monitor & Central Station.	Read as Point 21: Battery backup of minimum 1 hour when charged, in all high end monitor & Central Station.
6	Point 22: Facility for HIS Communications with information management systems includes laser printer and dual channel strip chart recorder.	Read as Point 22: Monitor should be compatible with installed pendant in BMHRC, Facility for Hospital Information System (HIS) Communications with information management systems includes laser printer and dual channel strip chart recorder.
7	Point 24: Accessories: (All above spares rate should be quoted separately)	Read as Point 24: Accessories: (All above spares rate should be quoted separately with item name)
8	Point 26: Mounting of Monitor on pendant	Read as Point 26: Monitor Should be compatible to Mount on exiting pendant
9	Point 29(3): Should be USFDA & CE approved product & it is mandatory	Read as Point 29(3): Should be USFDA & European CE approved product & it is mandatory
10	Point 17: Service facility should be Bhopal based.	Read as Point 17: Said point deleted


 Medical Superintendent