

भोपाल स्मारक अस्पताल एवं अनुसंधान केन्द्र

BHOPAL MEMORIAL HOSPITAL & RESEARCH CENTRE

स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार
Department of Health Research Ministry of Health & Family Welfare, Government of India
रायसेन बायपास रोड, भोपाल (म.प्र.) 462038 -Raisen By Pass Road, Bhopal – 462038,
Tel.: 0755-2742212 – 16 Fax: +91-755 274 83 09, Website: www.bmhrc.org

NO. BMHRC/PURCHASE/COORRIGEDUM/EQUIP/2017-18/350

30th Oct, 2017

CORRIGENDUM NOTICE

Ultra Sound Machine (USG 3D-4D COLOR DOPPLER)
Tender No. BMHRC/PUR/RE/2016-17/26 (3rd Call)

Reference above mentioned Tender and Corrigendum Notice NO. BMHRC/PURCHASE/COORRIGEDUM/EQUIP/2017-18/ 332 dated 20.10.2017. Further amendments pertaining to last date of Submission / Opening of above mentioned tender and modification in Technical Specification are as under:-

Last Date and time of submission of Tender Opening date and time of Technical- Bids as mentioned in the Corrigendum notice No.. BMHRC/PURCHASE/COORRIGEDUM/EQUIP/2017-18/ 332 dated 20.10.2017 and Technical Specification mentioned in Tender	Revised Last Date and Time of submission of Tender / Opening date and time of Technical – Bids and modification in Technical Specification of Tender
Last Date and Time of submission of Tender: 3 rd November, 2017 up to 11.00 a.m. Opening Date and time of Technical – Bids 3 rd November, 2017 at 12.00 Noon	Last Date and Time of submission of Tender: 17 th November, 2017 up to 11.00 a.m. Opening Date and time of Technical – Bids 17 th November, 2017 at 12.00 Noon
Point No. 9: (Page No. 13-15) The system shall have at least three universal transducer ports with electronic switching capability allowing any transducer to be connected to any port.	Point No. 9: (Page No. 13-15) Should be Read as: The system shall have at least four universal transducer ports with electronic switching capability allowing any transducer to be connected to any port.

All other terms and conditions mentioned in the Tender remain unchanged.

Note:- Any further modification / deletion / additions in the terms and condition of Tender or Technical Specification of the Equipment will be notified through Corrigendum Notice on the website of BMHRC or on CPP Portal.


Medical Superintendent
BMHRC, Bhopal