


CORRIGENDUM NOTICE

Tender No BMHRC/PUR/RE/2015-16/12 (Automated Elisa Processor)

S. No.	Particulars/ Technical Specification	Modifications in the Technical Specifications of the Tender [Please refer to Annexure 'A' Page No. 13]
1	Point 3(ii): Dilution unit with two precision syringes	Read as Point 3 (ii) Single probe with multi shot dispensing feature for dilution
2	Point 4(7): Mixer: 1300 rpm	Read as Point 4(7): Mixer: 1300 rpm: Mixing is available before taking the reading.
3	Point 4(12): Reading range: 0-3.000 O.D. monochromatic and bichromatic 0-9.000 O.D. polychromatic	Read as Point 4(12): Reading range: Reading range between 0.0 to 3.5 O.D with mono and bi-chromatic measurement


17/02/2012
Medical Superintendent
BMHRC, Bhopal