


CORRIGENDUM NOTICE

**Tender No BMHRC/PUR/LT/2016-17/21/RE.
Semi Auto Analyzer**

| S. No. | Particulars/ Technical Specification | Modifications in the Technical Specifications of the Tender [Please refer to Annexure 'A' Page No. 13] |
|---------------|--|--|
| 1 | Point 2: Detector: Photo diode -320 - 1000mm (6 filter 340-620) 6 optional position | Read as Point 2: Detector: Photo diode - 340 to 700 mm and optional position between 2 to 6. |
| 2 | Point 4: Calibration : Factors are point and multi point automatic on one standard (Linear model) automatic on up to 10 standard non linear mode | Read as Point 4: Calibration: Factors are point and multi point automatic on one standard (Linear model) automatic on up to 6- 10 standard non linear mode. |
| 3 | Point 9: Aspiration System: Aspiration Vol. programmable internal pump of bellow type | Read as Point 9: Aspiration System: Aspiration Vol. programmable internal pump |
| 4 | Point 11: Delay Time: Programmable 0 to 999 seconds | Read as Point 11: Delay time can be 5 sec to 999 sec |
| 5 | Point 12: Measuring Time: Programmable 2 to 998 seconds for kinetic and two point type of tests for end point fixed at 2 second | Read as Point 12: Measuring Time: Programmable 5 to 999 seconds for kinetic and two point type of tests for end point fixed at 2 second (approximately) |
| 6 | Point 13: Aspirational System: Internal bellows type of pump | Read as Point 13: Aspirational System: Peristaltic pump |


Medical Superintendent 17/10/2012
BMHRC, Bhopal