

भोपाल स्मारक अस्पताल एवं अनुसंधान केन्द्र
BHOPAL MEMORIAL HOSPITAL & RESEARCH CENTRE

स्वास्थ्यअनुसंधान विभाग, स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार
Department Of Health Research, Ministry Of Health And Family Welfare, Government Of India
रायसेन बायपास रोड, भोपाल - 462038 (म. प्र.) Raisen Bypass Road Bhopal - 462038 (M.P.)
Phones: +91 755 2742212-16, Fax: - +91 755 2748309, Email: bmhrcbhopal@gmail.com Website: WWW.bmhrc.org

CORRIGENDUM NOTICE

Subject: Corrigendum Notice for Tender No. BMHRC/PUR/2016-17/09 (Holium 100 Watt laser with Morcelletor and Flexible URS)

The following amendments in Technical Specification are approved by the Technical Specification committee:

S. No.	Particular /Technical Specification	Consideration of Committee Members
1	Point No. 6: It should have atleast two Laser Generators.	Read as Point No. 6: It should have two or above Laser Generators.
2	Point No. 7: It should have repetition rate of 5-50Hz.	Read as Point No. 7: It should have repetition rate in between 4 to 50Hz.
3	Point No. 8: It should have pulse duration up to 500 microseconds.	Read as Point No. 8: It should have pulse duration up to 500 or above microseconds.
4	Point No. 11: It should have a closed loop, self contained water to air exchanger cooling system.	Read as Point No. 11: It should have a closed loop, self contained water to air exchanger cooling system or active refrigeration cooling system
5	Point No. 14(c): 200 Micron Reusable, Flexible Fiber - 5	Read as Point No. 14(c) : 200 -275 Micron Reusable, Flexible Fiber - 5
6	Point No. 14(d): 550 Micron Side Fire Fiber for Ablation - 5	Read as Point No. 14(d) 550 Micron Side Fire Fiber for Ablation - 5 need one continuous flow laser cystoscope - 5
7	Additional Add New Point:	Read as Add Additional new point: Continuous Flow laser Cyst scope.



Medical Superintendent