

**भोपाल स्मारक अस्पताल एवं अनुसंधान केन्द्र**  
**BHOPAL MEMORIAL HOSPITAL & RESEARCH CENTRE**

स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
Department of Health Research Ministry of Health & Family Welfare, Government of India

NO. BMHRC/PURCHASE/CORRIGENDUM/EQUIP/2018-19/748

7<sup>th</sup> April, 2018  
10<sup>th</sup>

**CORRIGENDUM NOTICE**

**ECHOCARDIOGRAPHY MACHINE**  
**E-TENDER NO. BMHRC/PUR/2017-18/07 (2<sup>nd</sup> Call)**

Reference above mentioned E-Tender and minutes of Pre Bid Committee Meeting dated 6.04.2018. As per recommendation of pre bid committee following amendments /modification in technical Specification are as under:-

S. No.	Technical Specification as mentioned in Annexure –A of Tender (Page No. 18-19)	Amendments in Technical Specifications
1	<b>Point No. 8:-</b> Image storing facility on in built hard disc or MOD/CD/DV-RW facility should be available. In built- in hard disk the capacity should be of minimum 5 TB or 5,00,000 images. System should have extensive image management capability including thumbnail review, cine loop editing etc.	<b>Change as:-</b> Image storing facility on in built hard disc or MOD/CD/DV-RW facility should be available. In built-in hard disk the capacity should be of <b>minimum 1 TB</b> or 5,00,000 images. System should have extensive image management capability including thumbnail review, cine loop editing etc.
2	<b>Point No. 18:-</b> Should be provided with Adult Transthoracic 3 D probe, Paediatric Transthoracic 3 D probe, Adult Transesophageal 3 D probe, Paediatric Transesophageal 2D probe, Epi-aortic probe, vascular probe.	<b>Change as:-</b> Should be provided with Adult Transthoracic 3D probe, <b>Paediatric Transthoracic 2 D probe</b> , Adult Transesophageal 3 D probe, Paediatric Transesophageal 2D probe, Epi-aortic probe, vascular probe.

All other terms and conditions mentioned in the tender remain unchanged.

*Rama Desikan*

Director  
BMHRC, Bhopal